## PURCHASE DESCRIPTION

## SIGNAL GENERATOR (10 to 18 GHz)

## FSNTT-A

- 1.0 <u>GENERAL DESCRIPTION</u> This procurement requires a stable microwave signal generator capable of generating signals over the frequency range of 10 to 18 GHz with internal and external modulation capabilities.
- 2.0 <u>CLASSIFICATION</u> The equipment shall meet the requirements of MIL-T-28800(), Type III, Class 5, Style E, Color R for Navy shipboard, submarine, and shore applications with the following modifications and exceptions:
  - a. The non-operating temperature requirement is limited to the range of -40°C to +70°C.
  - b. The relative humidity requirement is limited to 95% noncondensating.
  - c. The operating and non-operating altitude requirements are not invoked.
  - d. The Electromagnetic Interference requirements of MIL-T-28800 are limited to CE01, CE03, CS01, CS02 (0.05 to 100 MHz), CS06, RE01 (back panel search excluded), RE02 (14kHz to 1 GHz), and RS03.
  - e. The warm-up time is extended to one hour.
- 3.0 <u>OPERATIONAL REQUIREMENTS</u> The equipment shall be capable of generating signals within the parameters and accuracies specified herein.
- 3.1 <u>Frequency Characteristics</u>
- 3.1.1 Frequency Range: At least 10 to 18 GHz
- 3.1.2 Frequency Resolution: Minimum resolution at least 1 kHz; digital readout
- 3.1.3 Frequency Accuracy: Indicated frequency shall be within ±1 kHz plus the measured offset the time base when measured on an external counter.
- 3.1.4 Frequency Stability (equal to or better than limits specified below)
- 3.1.4.1 Internal: Better than 1 part in 109/hour after one hour warm-up
- 3.1.4.2 External: Equal to external standard frequency stability
- 3.1.5 Residual FM (CW mode in 50 Hz to 15 kHz detection BW): Less than 1.0 kHz peak
- 3.1.6 Spectral Purity (equal to or better than limits specified below)
- 3.1.6.1 Harmonics: All harmonically related outputs shall be at least -55 dBc.

3.1.6.2 3.1.6.3

3.2	Output Characteristics
3.2.1	Range: +10 to -110 dBm (minimum)
3.2.2	RF Output: Leveled output shall be at least +10 dBm
3.2.3	Accuracy: ±2.0 dB of the actual measured output level, +10 to -110 dBm
3.2.4	Display/Resolution: Digital display; minimum resolution of 0.1 dB
3.2.5	Flatness: ±1.0 dB measured at an output level of +10 dBm
3.2.6	Impedance/Connector: 50 ohms; type N female connector
3.2.6.1	VSWR: 2:1 maximum
3.2.7	Reverse Power Protection: The generator shall be capable of accepting the following signal levels at its output connector without resulting damage.
3.2.7.1 3.2.7.2	Average Power: 4 W Peak Power: 3.5 kW peak for pulse widths of less than 1 microsecond
3.3	Modulation Characteristics
3.3.1	Pulse Modulation
3.3.1.1 3.3.1.1.1 3.3.1.1.2 3.3.1.1.3 3.3.1.1.4 3.3.1.1.5	Internal Pulse Rate (PRF): At least 100 Hz to 50 kHz Pulse Width (PW): 0.1 to 10.0 microseconds Rise and Fall Times: Less than 50 nanoseconds ON/OFF Ratio: Greater than 80 dB Delay: At least 100 nanoseconds to 10 milliseconds; accuracy 20% of setting
3.3.1.1.5. 3.3.1.1.5. 3.3.1.2	
3.3.1.2.1 3.3.1.2.2	Pulse Rate: At least 100 Hz to 50 kHz Pulse Width: Greater than 0.1 microseconds
3.3.2	Amplitude Modulation (AM)
3.3.2.1 3.3.2.1.1 3.3.2.1.2	Internal AM (square wave) Rate: At least 1 kHz Depth: At least 0 to 70%

Non-harmonics/Spurious: At least -55 dBc at frequencies greater than 1 kHz from carrier Phase Noise: At least -75 dBc/Hz at 10 kHz offset from carrier

3.3.2.2.1 3.3.2.2.2 3.3.2.2.3	· ·
3.3.3	Frequency Modulation (FM)
3.3.3.1	Internal FM
3.3.3.1.1 3.3.3.1.2	Rate: At least 100 Hz to 50 kHz FM Deviation: At least 1 MHz peak locked or unlocked; unlock center frequency shift
	less than 50 MHz
3.3.3.2	External FM
3.3.3.2.1 3.3.3.2.2	Rates: At least 100 Hz to 50 kHz FM Deviation: At least 1 MHz peak locked or unlocked; unlock center frequency shift less than 50 MHz
4.0	GENERAL REQUIREMENTS
4.1	<u>Power Source</u> : 115 or 230 Vac $\pm$ 10%, single phase, 50, 60 or 400 Hz $\pm$ 10%, 200 W maximum
4.2	<u>Calibration Interval</u> : The calibration interval shall be 12 months minimum. The equipment shall be within all accuracy requirements specified herein, with a 72% or greater confidence factor following a calibration interval of 12 months.
4.3	<u>Dimensions</u> : The total volume of the unit shall not exceed 1800 cubic in (29,500 cubic cm).
4.4	Weight: The overall weight of the unit shall not exceed 50 pounds (22.7 kg).
4.5	Remote Operation: The unit will be capable of remote operation via IEEE-488 bus interface. It shall operate as a talker or listener such that all functions except the power on/off switch are

controllable and shall have as a minimum the following subset of GPIB commands: AH1, SH1,

T6, L4, SR1, RL1, DC1, DT1.